

Analyzer Pressure Regulation and Vent Recovery System

Bulletin 4141-VR November 2001



Model Shown: VRAZ3-01-22-DX-V-SS

Unparalleled pressure and flow stability

The Parker Analyzer Pressure Regulation and Vent Recovery System automatically and continuously adjusts for variations in gas supply pressures and flows. This system has been engineered to meet the requirement of continuous analyzers designed to operate at constant pressure in order to provide accurate analyses of gases. The combination of pressure and flow regulation provides the required stability even with greatly varying inlet and outlet pressures.

Features

- Designed for a zero, a calibration, and up to two sample streams
- Capabilities Stream switching Sample filtering Flow and pressure regulation
- Regulator controls analyzer pressure
- Metering valves control Analyzer flow Stream by-pass flow
- Flowmeters indicate stability of Analyzer flow
 Stream by-pass flow
- Pressure gauges indicate Sample inlet pressure Pressure to the analyzer Pressure from the filter
- Utilizes Parker
 - *R-Max*[™] Stream Switching System MB Series Ball Valves HR Series Metering Valves IR5000 Pressure Regulator SC Flow Controller Balston[®] Particulate Filtration CPI[™] Connectors
- 100% Factory Tested
- Patent Pending
- Custom Engineered Systems Available

Operating Conditions

- Pressure Rating: 200 psig (14 bar) maximum
 Temperature Rating:
- Media -40 °F to 200 °F (-40 °C to 93 °C) Ambient -40 °F to 140 °F (-40 °C to 60 °C)

Functional Performance

 Pressure Ratings: IR5000 Pressure Regulator 3500 psig (241 bar) MB Series Ball Valve 3000 psig (207 bar) *R-Max[™]* Stream Switching System 500 psig (34 bar) HR Series Metering Valve 250 psig (17 bar) Flowmeter 200 psig (14 bar) SC Flow Controller 150 psig (10 bar)

Materials of Construction*

Wetted

Wetted	
<i>R-Max</i> [™] System	
Base, Body and Stems	316 Stainless Steel
Upper and Lower Seats	
Seals	
IR5000 Regulator	
Body	3161 Stainless Steel
Diaphragm	
	-
Compression Member	
Poppet	
Poppet Spring	
Carrier	
Seat	PCTFE
Back-up O-ring	Fluorocarbon Rubber
Inlet Screen/Filter	316L Stainless Steel
MB Ball Valves	
Body and Stem	316 Stainless Steel
Seat/Packing	
HR Metering Valves	
Cartridge Components	216 Staiplace Staal
Orifice Liner	
Stem Seals	Fluorocarbon Rubber
SC Controller	
Body and Piston	
Seat and Seals	Fluorocarbon Rubber
Diaphragm	Hastelloy C-22 [®]
Flowmeters	
Body	316 Stainless Steel
Tube	Borosilicate Glass
Float	Glass
Float Stops	
Gauges	
Body	216 Staiplace Staal
Bourdon Tube	
Fittings	
Tubing	316 Stainless Steel
Non-wetted	
Panel	304 Stainless Steel
Brackets	
Hardware	
Operating Handles ABS Plastic, Nylon 6/6	
Gauge and Flowmeter Shields Polycarbonate	
* Materials of construction for stainless steel systems.	
Consult factory for optional materials of construction.	
Hastellov C-22 is a registered trademark of Havnes International. Inc.	

Hastelloy C-22 is a registered trademark of Haynes International, Inc. Inconel is a registered trademark of Inco Alloys International Elgiloy is a registered trademark of Elgiloy Company

© Copyright 2001, Parker Hannifin Corporation. All Rights Reserved.





Performance Example 1 - Varying Inlet Pressure with Constant Outlet Pressure



Analyzer Pressure set to 5 psig (0.3 bar); Supply Pressure varied from 20 to 90 psig (1.4 to 6.2 bar)

Performance Example 2 - Varying Outlet Pressure with Constant Inlet Pressure



Analyzer Pressure set to 12.7 psia (0.87 bar); Supply Pressure set to 14.5 psia (1.01 bar); Outlet Pressure varied from 1 to 11 psia (0.07 to 0.76 bar)

Options

By-pass Filter - Approximately 90% of the inlet flow by-passes the cartridge filter and exists the filter bowl. This provides three benefits: 1) It reduces the transport time of the sample stream from the process line to the analyzer; 2) Provides a continuous flushing action on the filter element; and, 3) The life of the filter element is greatly extended since only a small percentage of the flow is filtered - and only when the stream is selected for analysis. **Manual Stream Switching** - The *R-Max*[™] Stream Switching System is replaced with two-way and three-way MB Series manual Ball Valves. The optional By-pass Filter(s), if ordered, is mounted as a stand-alone unit downstream of the sample inlet MB Series manual Ball Valve(s).

Integral Aspirator - A Parker VC Vacuum Generator is added between the Flow Controller and MB Series Ball Valve on the Sample Return line. A Parker IR4000 Pressure Regulator is also added to control the vacuum generated.

Filter Purge Valve - A three-way MB Series manual Ball Valve is placed upstream of the Filter to enable switching between sample and purge gas.

Aspirator and Filter Purge Valve - Adds both the Integral Aspirator and the Filter Purge components to the panel.

Analyzer Pressure Regulation and Vent Recovery System



Dimensions**



() Denotes dimensions in millimeters



Parker Hannifin Corporation

Instrumentation Valve Division 2651 Alabama Highway 21 North Jacksonville, AL 36265-9681 Phone: (256) 435-2130 Fax: (256) 435-7718 www.parker.com/IVD